

Model : AM-P-PH Resolution : 0.1

 Range: 0-14 Accuracy: ±2%

Battery: 2 x LR44 Cell

Weight: 55g

# Pen Type pH Meter

Kindly inspect the equipment for any signs of damage and verify that you have received functional equipments.

### Calibration

- Remove the protection cap from meter to rinse the electrode with clean water and wipe it dry and replace protective cap at end of usage.
- Dip the electrode into the buffer solution pH 7.0 and stir gently until the reading is stable. To Calibrate Press and hold "CAL" for 5 Sec. Display will flash and return to measuring mode.
- 3. Similarly to calibrate for pH 4.0 and pH 10.0, Take respective standard solutions and follow Step 2.

### Note

- Calibration is not necessary before each usage, but it should be done every month.
- 2. Don't touch or wipe the surface of inner sensor
- 3. Change a new battery when the display fades or flashes.

### Measurement

- 1. After calibration, rinse the electrode with clean water and wipe it dry. Dip the electrode into sample solution to be measured. Stir gently and wait until a stable reading can be obtained.
- 2. After measurement, rinse the electrode with clean water and replace protective cap.

# Maintenance

Battery replacement

- Pull out the battery compartment.
- 2. Remove the old Lithium battery LR44 and note
- Replace the battery compartment press fit.

## Safety:

- Use the instrument under professional instructions to avoid failure
- 2. Avoid Heat or direct Sunlight
- 3. Use standard buffer solution for calibration to avoid false readings
- Do not insert any other part of the instrument in solution apart from the sensor to avoid circuit failure.